



Product Service

CERTIFICATE

No. Z1 17 07 22430 049

Holder of Certificate: Alu Rehab ASBedriftsvegen 23
4353 Klepp Stasjon
NORWAY**Production Facility(ies):** 57370**Certification Mark:****Product:** Standard wheelchairs
Manual wheelchairs**Model(s):** Netti 4U CES PLUS
Netti 4U CE PLUS**Parameters:**
total length incl. footrest:
Netti 4U CES Plus Netti 4U CE Plus
950 - 1050 mm 960 - 1160 mm

seat depth:
Netti 4U CES Plus Netti 4U CE Plus
375 - 450 mm 425 - 500 mm

Remark: Further parameters see enclosure

Tested according to: DIN EN 12183:2014

The product was tested on a voluntary basis and complies with the essential requirements. The certification mark shown above can be affixed on the product. It is not permitted to alter the certification mark in any way. In addition the certification holder must not transfer the certificate to third parties. See also notes overleaf.

Test report no.: 713096374**Valid until:** 2022-07-09**Date,** 2017-07-10
(Torsten Zimmer)

Page 1 of 2



Enclosure to certificate no.: Z1 17 07 22430 049

	Netti 4U CES Plus	Netti 4U CE Plus
width:	530 - 680 mm	
height:	1100 mm (w/o head rest)	
seat width:	350 - 500 mm	
seat height:	465 - 500 mm	
backrest height:	500 mm	
Seat plane angle:	-5° - 20°	
backrest angle:	90° - 135°	
max. user weight:	135 kg	
brake system:	knee lever brake (user brake)	
back wheels:		
-size	22 x 1", 24 x 1" and 24 x 1 3/8"	
-tyres	solid rubber (puncture proof)	
-adjustable	yes	
-fixation	quick release	
front wheels:		
-size	ø150 x 27 and ø175 x 33 mm	
-tyres	solid rubber (puncture proof)	
-adjustable	yes	
-fixation	quick release	
frame material:	aluminum	
armrest:	height adjustable and removable	
legrest:	height adjustable, angle adjustable and removable	
anti tip device:	adjustable	

Trade Mark:



Department: MHS5 / me
Date: 2017-07-09

