



Product Service

CERTIFICATE
No. Z1 10 11 22430 026

Model(s): **Manual wheelchair type: Netti II**

Parameters:

total length	
incl. footrests:	1100 mm
total length	
without footrests:	900 mm
width:	590 - 740 mm
height:	1120 mm
seat width:	350, 380, 400, 430, 450 and 500 mm
seat depth:	400 - 500 mm
seat height:	400 - 480 mm
backrest height:	480 mm (600 mm with back extension)
seat plane angle:	- 9° - +16°
backrest angle:	83° - 107°
max. user weight:	125 kg
brake system:	knee lever brake (user brake)
back wheels:	
- size	20 x 1", 22 x 1", 24 x 1" and 24 x 1 3/8"
- tyres	solid rubber (accident proof)
- adjustable	yes
- fixation	Quick release
front wheels:	
- size	ø 125 x 27, ø 150 x 27 and 175 x 33 mm
- tyres	solid rubber (accident proof)
- adjustable	yes
- fixation	Quick release
frame material:	aluminium, steel
armrest:	height adjustable and removable
legrest:	height adjustable, angle adjustable and removable
anti tip device:	adjustable

Tested according to: DIN EN 12183:2009

Production Facility(ies): 22430



Product Service

CERTIFICATE

No. Z1 10 11 22430 026

Holder of Certificate: **Alu Rehab AS**

Bedriftsveien 23
4353 Klepp Stasjon
NORWAY

Certification Mark:



Product:

**Standard wheelchairs
Manual wheelchairs**

The product was tested on a voluntary basis and complies with the essential requirements. The certification mark shown above can be affixed on the product. It is not permitted to alter the certification mark in any way. In addition the certification holder must not transfer the certificate to third parties. See also notes overleaf.

Test report no.:

71375625



Date, 2010-11-25

(Torsten Zimmer)

Page 1 of 2